

<i>ênica</i>	<i>Valores da mediana encontrados (dB)</i>	<i>ISO 8253-1 Padrão V (dB)</i>	<i>Diferença entre Padrão V e valores da mediana (dB)</i>	<i>ISO 8253-1 Padrão X (dB)</i>	<i>Diferença entre Padrão V e valores da mediana (dB)</i>
z)	61,0	63,0	-3,0	71	- 10,0
,5	45,5	44,0	+1,5	52	- 6,5
3	32,0	28,0	+ 4,0	36	- 4,0
15	30,0	13,0	+ 17,0	21	+ 9,0
30	28,5	8,0	+ 20,5	16	+ 8,0
60	30,0	7,0	+ 23,0	15	+ 7,0
90	27,0	8,0	+ 19,0	16	+ 11,0
120	26,5	2,0	+ 24,5	10	+ 16,5
150	26,5	15,0	+ 11,5	23	+ 3,5